

If the AutoTest Emissions Tester installer shows the following message: "The wizard was interrupted before AutoTest Emissions Tester could be completely installed."

1. Go to the Program Files directory on the C Drive
2. Find the AutoTest folder
3. Right click on the folder
4. Click on the "Security" tab
5. For your user account which should be under your name or administrator click on the "Edit" button
6. Under the "permissions for everyone" list, tick "Allow" for "full control" and "modify"
7. Click Apply and ok
8. The installer should now work

How often does the Gas Analyser need to be calibrated?

Every 12 months. Send it directly to us, or via your local distributor for recalibration.

Can the Gas Analyser connect to the computer using USB?

The Gas Analyser needs to be connected to the PC via the supplied cable. Earlier cables are RS232 and will need an RS232 to USB adapter to plug into a USB port. Current cables connect directly through the USB port to the Gas Analyser.

Where is the user manual?

The user manual is stored inside the USB that came in the box. The manual can also be downloaded from our website.

Where do I buy the paper?

You can purchase a paper pack through our website or your local distributor. A paper pack consists of 4 paper rolls.

Can I buy more filters?

Yes. You can purchase these directly through our website.

Can I replace sensors by myself?

No. The Gas Analyser needs to be recalibrated after changing any sensors. Please contact AutoTest Products.